


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> CALIBRATION</p> <p><b>9460</b></p> <p>Accredited to <b>ISO/IEC 17025:2017</b></p>	<h3>NDT Maincal Limited</h3> <p>Issue No: 013 Issue date: 19 May 2023</p>	
	<p><b>Unit 1a</b> Bingswood Trading Estate Whaley Bridge High Peak SK23 7LY</p>	<p><b>Contact: Lee Wilde</b> Tel: +44 (0) 1663 735283 Fax: +44 (0) 1663 733482 E-Mail: Lee@maincal.com Website: www.maincal.com</p>
<p><b>Calibration performed by the Organisation at the locations specified</b></p>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<p><b>Address</b> NDT Maincal Limited Unit 1a Bingswood Trading Estate Whaley Bridge High Peak SK23 7LY United Kingdom</p> <p><b>Local contact</b> Lee Wilde Tel: +44 (0) 1663 735283 Fax: +44 (0) 1663 733482 E-Mail: Lee@maincal.com</p>	<p>Magnetic particle inspection and associated equipment</p> <p>Ultrasonic test equipment</p> <p>Pressure indicator</p> <p>AC conductivity</p>	Lab

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p>The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.</p>	<p>Magnetic particle inspection and associated equipment</p> <p>Ultrasonic test equipment</p> <p>Pressure indicator</p> <p>Temperature</p>	Site



9460

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**NDT Maincal Limited**  
**Issue No:** 013    **Issue date:** 19 May 2023

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
ELECTRICAL				
All electrical measurements a carried out using the method of direct comparison or transfer to laboratory reference standards unless otherwise determined in the remarks column.				
DC Current	1 mA to 2 A	0.80 %		Lab & Site
Voltage Peak to Peak	2 A to 20 A	2.2 %		
10 Hz to 20 MHz	0.6 Vpk-pk to 40 Vpk-pk	3.0 %		Lab & Site
	40 Vpk-pk to 390 Vpk-pk	4.2 %		
DC Current	20 A to 50 A	5.0 %		Lab & Site
All waveforms	50 A to 3 kA	3.0 %		
AC Voltage				Lab & Site
RMS 50 Hz	100 V to 450 V	2.5 %		
AC Current				Lab & Site
50 Hz	1 mA to 2 A	1.4 %		
	2 A to 20 A	2.6 %		
Attenuation				Lab & Site
10 Hz to 25 MHz	0 dB to 90 dB	.050 dB		
Nominal input 5 V to 10 V	90 dB to 110 dB	0.10 dB		
Frequency	10 Hz to 100 MHz	0.50 %		Lab & Site
HWDC scaled in MAGAMPS (2XDC average)	100 A to 6 kA	3.0 %	Ranged is scaled up by 2	
50 Hz rms	50 A to 3.5 kA	3.0%		Lab & Site
50 Hz peak	50 A to 4.95 kA	3.0 %		
AC & DC half wave peak current	50 A to 2.5 kA	4.2 %		Lab & Site
	2.5 kA to 7.0 kA	4.5 %		
Current Shot elapsed time	0 s to 1.9 s	22 ms		Lab & Site
	1.9 s to 4.8 s	56 ms		



9460

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**NDT Maincal Limited**  
**Issue No: 013 Issue date: 19 May 2023**

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
AC Conductivity Nominal 60 kHz Sourcing nominal values	2 MS/m to 60 MS/m	0.75 %	Note; 58.0 MS/m = 100 % on the International Annealed Copper Scale	Lab
Measurement	2 MS/m to 60 MS/m	1.3%		
Traceability to Boeing (NIST)	0.5 MS/m to 36 MS/m 36 MS/m to 60.5 MS/m	0.75 % 1.5 %	Boeing BAC 5651 Table III. Results will be reported using the IACS Scale, the conversion to SI units will be reported on the certificate.	
Thickness	0 mm to 100 mm	0.50 mm	Supplementary information for the calibration of conductivity blocks only	
ELECTRICAL VERIFICATION of ULTRASONIC FLAW DETECTION EQUIPMENT	As BS EN 12668-1:2010 Group 2 tests and including the following calibrations and quantities:  Stability after warm up (height) Stability after warm up (width) Jitter - screen height Jitter - screen width  Stability against voltage variation (height) Stability against voltage variation (width)  Transmitter voltage 50 V to 500 V  Pulse risetime 30 V to 450 V  Pulse duration 1 ms to 25 ns  Pulse Reverberation	  0.50 % of screen height 0.50 % of screen width 0.50 % of screen height 0.50 % of screen width 0.50 % of screen height 0.50 % of screen height  0.5 % of screen width  4.2 %  3.0 ns  2.5 ns + 2.0 %  2.5 V	The method of measurement is described in the standards listed.  These claims are all dominated by the resolution of the digital readout rather than the display resolution.	Lab & Site



9460

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**NDT Maincal Limited**  
**Issue No:** 013    **Issue date:** 19 May 2023

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
ELECTRICAL VERIFICATION of ULTRASONIC FLAW DETECTION EQUIPMENT continued	Amplifier frequency response 100 kHz to 50 MHz	4.0 % of screen height	The method of measurement is described in the standards listed.	Lab & Site
	Equivalent input noise	5.0 nV/ $\sqrt{\text{Hz}}$		
	Linearity of vertical display	2.5 % of screen height		
	Linearity of timebase x axis	1.0 % of screen width		
	As BS EN ISO 22232-1:2020 Group 2 tests and including the following calibrations and quantities:		Method B	
	Transmitter voltage 50 V to 500 V	4.2 %		
	Pulse risetime 30 V to 450 V Pulse duration 1 ms to 25 ns	3.0 ns 2.5 ns + 2.0 %		
Amplifier frequency response 100 kHz to 50 MHz Equivalent input noise	4.0 % of screen height 5.0 nV/ $\sqrt{\text{Hz}}$			
Linearity of vertical display	2.5 % of screen height			
Eddy Current measurements Attenuation, probe drive voltage and frequency are listed in the general electrical capabilities.			ISO 15548-1:2013 tests as listed.	Lab & Site
Screen angle rotation	0° to 360° in 45° degree increments.	1.0°	Correlation between UUT digital readout and display.	Lab & Site
Overall system measurements with probe.	0.20 mm, 0.50 mm and 1.0 mm	3.0 % of full screen width	Serial number of probe used will be stated on the certificate	Lab & Site
Magnetics measurements	2 Gauss to 30 Gauss	0.50 Gauss		Lab



Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**NDT Maincal Limited**  
**Issue No:** 013    **Issue date:** 19 May 2023

**Calibration performed by the Organisation at the locations specified**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
<b>PRESSURE</b>  Gas pressure gauge  Calibration of pressure indicating instruments and gauges	0 Pa to 200 kPa 200 kPa to 400 kPa 400 kPa to 700 kPa	2.3 kPa 3.6 kPa 4.2 kPa	Calibration by comparison with a digital pressure standard. Calibrations can be undertaken in other pressure units.	Lab & Site
<b>Temperature</b>  Temperature controlled ovens and associated indicators  Water temperature indicators	45 °C to 75 °C  1 °C to 55 °C	1.5 °C  2.5 °C	Single and multipoint monitoring probes Time dependent temperature profiling	Site  Lab & site
<b>END</b>				



Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**NDT Maincal Limited**  
**Issue No:** 013    **Issue date:** 19 May 2023

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$